




ELK Biotechnology
For research use only.

DNA Marker II

Catalog No.	Specification	Storage/Shelflife
EDL110-01	250 μ l	-20C°/2 years
EDL110-02	500 μ l	-20C°/2 years

Introduction

DNA Marker II is composed of 6 kinds of DNA fragments with a length of 100 bp to 1,200 bp, dissolved in 1 \times Loading Buffer, and can be used for direct electrophoresis of 5-10 μ l during used, which is very convenient to use.



Note

1. When the width of the sample well during electrophoresis is less than 5 mm, a clear band can be obtained by taking 5 μ l Marker electrophoresis each time. If the sample hole is widened, the sampler volume of Marker must be increased appropriately.
2. For DNA electrophoresis, the purity of Agarose greatly affects the clarity of DNA bands. Therefore, the best quality Agarose should be used for electrophoresis, and the recommended gel concentration is 2% to 3%.
3. When performing Agarose electrophoresis, the concentration of Agarose is closely related to the separation performance of DNA fragments. The greater the concentration of Agarose, the better the separation performance of short DNA; conversely, the smaller the concentration of Agarose, the more favorable for the separation of long DNA.